

AmulFed Dairy

A UNIT OF GUJARAT CO-OPERATIVE MILK MARKETING FEDERATION LIMITED

(PREVIOUSLY KNOWN AS MOTHER DAIRY, GANDHINAGAR)

PURCHASE ENQUIRY

Inquiry No	: AFD/PUR/PKS/ENQ/PI, Dated: Sep 11, 2018
Description	: Purchase Inquiry for Supply, installation and commissioning of Electronics System Weighing Scale With Ethernet Interface(LAN), Alarm Buzzer, Acrylic Cabinet Box and SS standards weights
Qty	: As per Annexure -I.
Due Date	: 25 th Sep, 2018

Dear Sir,

You are requested to give your most competitive rates for Supply, Installation and Commissioning of Electronics System Weighing Scale With Ethernet Interface(LAN), Alarm Buzzer, Acrylic Cabinet Box and SS standards weights at AmulFed Dairy at Gandhinagar with all necessary accessories required, as per the annexure attached herewith:

The offer should reach, AmulFed Dairy (A Unit Of Gujarat Co - Operative Milk Marketing Federation Limited.) Plot No. - 35, Near Indira Bridge, Ahmedabad - Gandhinagar Highway, Village: Bhat, Dist.: Gandhinagar. Pin No.382 428.

- The offer should be sent in sealed cover super scribing on the top left hand corner offer for inquiry No. AFD/PUR/PKS/ENQ/PI dtd. 11/09/2018. Offer received after due date for whatsoever reason may be rejected.
- Your offer should be sent through Post/Courier/mail & not by Fax or personally.
- Offer should be valid for acceptance, for 60 days from due date of submission of offer.
- The price should be quoted on F.O.R basis inclusive of packing and forwarding, taxes, excise duties, freight etc. However if it is separate, it should be clearly mentioned in the quotation as per the attached annexure.
- It should cover a warranty of at least one-year against manufacturing defect from commissioning of the plant.
- Realistic earliest delivery period should be indicated.
- It is not binding on us to accept lowest offer.
- Right to reject any or all the offers without assigning any reason is reserved by us.
- Offer which are incomplete or not meeting the conditions or only for part items, are liable for rejections.
- No queries regarding the status of the offer will be entertained at any time.
- The final supplier will be the combined decision of our technical team and management and no reason of final selection will be conveyed to any bidder.
- Payment terms with 30 days credit shall be preferred.
- We are an ISO 14001 certified unit. We follow Eco- friendly processes in our factory & also expect same to be followed in your company.
- Contact detail: Mr. Palak Shah, Mail: palak.shah@amul.coop, Mob: 7359001666

Amul SAGAR

Plot No. 35, Nr. India Bridge, Ahmedabad – Gandhinagar Highway, Village: Bhat, Dist.: Gandhinagar, PIN 382 428

Tel. No.: (079) 23969055 – 58, Fax No.: (079) 23969059

Website: www.amul.com

AmulFed Dairy

A UNIT OF GUJARAT CO-OPERATIVE MILK MARKETING FEDERATION LIMITED

(PREVIOUSLY KNOWN AS MOTHER DAIRY, GANDHINAGAR)

Annexure-I

Sr No.	Instrument	No. of Qty (EA)	Technical specification
1	Electronics System Weighing Scale With Ethernet Interface(LAN), Alarm Buzzer, Acrylic Cabinet Box and SS standards weights	02	<p>Technical specifications are as under:-</p> <p>Capacity : 3 Kg</p> <p>Accuracy : 0.2 gm</p> <p>Platter : SS -304</p> <p>Framework : SS-304</p> <p>32-bit ARM processor and 24-bit precision delta-sigma ADC</p> <p>Selectable filter and stability settings for any weighing application</p> <p>Automatic self-diagnostic function for easy trouble shooting</p> <p>Wide variety of configurable printing formats and reports</p> <p>Cloning the scale configuration and data for faster and reliable replacement</p> <p>Compatible ERP scale server</p> <p>User-friendly interactive menus</p> <p>Secure access control through passwords in software</p> <p>Optional second platform and external scale/balance interface</p> <p>Windows based PC application software (Compatible for Serial & Ethernet port)</p> <p>Functionality : PLU manager,File manager, Field manager, Scale management. etc.,</p> <p>Basic weighing/Counting version : OS Support Windows XP/7/8</p> <p>Filling version : OS Support Windows XP</p> <p>Run time of report should be in real time, and all entry should be in auto in excel file .</p> <p>Flash memory should be 8 MB</p> <p>Minimum 30 character for sample ID with combination of alpha, numeric, symbol.</p> <p>Tower lamp with Buzzer</p> <p>real time weigh balance display should be show on computer monitor</p> <p>standard stamping for at least 01 year for new supply weigh balance along with standard appropriate recommended SS weight for verification (e.g, 200g, 500g, 1 kg etc.)</p> <p>Centronics parallel</p> <p>RS-485 (MODBUS ASCII)</p> <p>Digital I/O</p> <p>USB Host -2 Nos.</p> <p>Loadcell connector</p> <p>Current or voltage interface for PLC</p>

Amul SAGAR

Plot No. 35, Nr. India Bridge, Ahmedabad – Gandhinagar Highway, Village: Bhat, Dist.: Gandhinagar, PIN 382 428

Tel. No.: (079) 23969055 – 58, Fax No.: (079) 23969059

Website: www.amul.com

AmulFed Dairy

A UNIT OF GUJARAT CO-OPERATIVE MILK MARKETING FEDERATION LIMITED

(PREVIOUSLY KNOWN AS MOTHER DAIRY, GANDHINAGAR)

			keyboard interface -1 Nos	
			barcode scanner interface -Nos.	
			Ethernet interface	
			RS-232 interface -2Nos.	
			Potential free set point relay output	
			Power on/off switch	
			Power cord	
			Limit indicators for error free operation	
			Unit selection (kg, g, ct, lt.,)	
			Polyester key sheet for long durability	
			Multi-tap alpha-numeric key board	
			High resolution graphic display for easy readability under any condition	
			Exhaustive PLU function management	
			Durable heavy duty keys for weigh acknowledge	
			Standard check weighing / count function	
			Printing data: print (for each transaction) & List print (summary reports)	
			Report with Statistical analysis and summerised details(Product Name, Date & Time, Tare wt., Net wt., Gross wt., Avg Net Wt, Lowest Wt, Highest wt, Std Dev, Variance, Rel. Deviation, No. Of Low, No Of OK, No Of High wt etc.)	
2	Check Weigher	1	Check Weigher-6KG Accuracy-0.5GM Plat Form-SS With USB option Indented Weighing scale is competent with USB option. We can Store weighment data in Pen drive and copy it to PC day wise/shift wise.	

Note:

Above is the General requirement for the instruments.

In case of any emergency breakdown or maintenance the service engineer should be available within 24-48 hours for repairing .The spares should be arranged within 24-48 hours during the entire life of instrument.

Amul SAGAR

Plot No. 35, Nr. India Bridge, Ahmedabad – Gandhinagar Highway, Village: Bhat, Dist.: Gandhinagar, PIN 382 428

Tel. No.: (079) 23969055 – 58, Fax No.: (079) 23969059

Website: www.amul.com

AmulFed Dairy

A UNIT OF GUJARAT CO-OPERATIVE MILK MARKETING FEDERATION LIMITED

(PREVIOUSLY KNOWN AS MOTHER DAIRY, GANDHINAGAR)

Amul SAGAR

Plot No. 35, Nr. India Bridge, Ahmedabad – Gandhinagar Highway, Village: Bhat, Dist.: Gandhinagar, PIN 382 428

Tel. No.: (079) 23969055 – 58, Fax No.: (079) 23969059

Website: www.amul.com